

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

**Client:** Aires Consulting Group  
**Project:** Berkeley District 87/ Sunnyside School  
**Client Sample ID:** SWF-1A  
**Collection Date:** 11/16/17 5:23

**Lab Order:** 17K0763  
**Lab ID:** 17K0763-01  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:24	12/7/17 21:01	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-1B  
**Collection Date:** 11/16/17 5:23

**Lab ID:** 17K0763-02  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:24	12/7/17 21:04	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-2A  
**Collection Date:** 11/16/17 5:24

**Lab ID:** 17K0763-03  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 21:24	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-2B  
**Collection Date:** 11/16/17 5:24

**Lab ID:** 17K0763-04  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 21:33	EPA200.8 R5.	JTC

**Client Sample ID:** SS-3A  
**Collection Date:** 11/16/17 5:27

**Lab ID:** 17K0763-05  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 21:36	EPA200.8 R5.	JTC

**Client Sample ID:** SS-3B  
**Collection Date:** 11/16/17 5:27

**Lab ID:** 17K0763-06  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 21:39	EPA200.8 R5.	JTC

**Client Sample ID:** SS-4A  
**Collection Date:** 11/16/17 5:28

**Lab ID:** 17K0763-07  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 21:42	EPA200.8 R5.	JTC



Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Client:</b> Aires Consulting Group <b>Project:</b> Berkeley District 87/ Sunnyside School <b>Client Sample ID:</b> SCS-7B <b>Collection Date:</b> 11/16/17 5:37 <b>Lab Order:</b> 17K0763 <b>Lab ID:</b> 17K0763-14 <b>Matrix:</b> Drinking Water									
<b>Client Sample ID:</b> SWF-8A <b>Collection Date:</b> 11/16/17 5:39 <b>Lab ID:</b> 17K0763-15 <b>Matrix:</b> Drinking Water									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:20	EPA200.8 R5.	JTC
<b>Client Sample ID:</b> SWF-8B <b>Collection Date:</b> 11/16/17 5:39 <b>Lab ID:</b> 17K0763-16 <b>Matrix:</b> Drinking Water									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:22	EPA200.8 R5.	JTC
<b>Client Sample ID:</b> SS-9A <b>Collection Date:</b> 11/16/17 5:42 <b>Lab ID:</b> 17K0763-17 <b>Matrix:</b> Drinking Water									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:34	EPA200.8 R5.	JTC
<b>Client Sample ID:</b> SS-9B <b>Collection Date:</b> 11/16/17 5:42 <b>Lab ID:</b> 17K0763-18 <b>Matrix:</b> Drinking Water									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:37	EPA200.8 R5.	JTC
<b>Client Sample ID:</b> SWF-10A <b>Collection Date:</b> 11/16/17 5:55 <b>Lab ID:</b> 17K0763-19 <b>Matrix:</b> Drinking Water									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:40	EPA200.8 R5.	JTC
<b>Client Sample ID:</b> SWF-10B <b>Collection Date:</b> 11/16/17 5:55 <b>Lab ID:</b> 17K0763-20 <b>Matrix:</b> Drinking Water									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:43	EPA200.8 R5.	JTC

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

**Client:** Aires Consulting Group  
**Project:** Berkeley District 87/ Sunnyside School  
**Client Sample ID:** SWF-11A  
**Collection Date:** 11/16/17 5:47

**Lab Order:** 17K0763  
**Lab ID:** 17K0763-21  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:46	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-11B  
**Collection Date:** 11/16/17 5:47

**Lab ID:** 17K0763-22  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:27	12/7/17 22:49	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-12A  
**Collection Date:** 11/16/17 5:49

**Lab ID:** 17K0763-23  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 22:57	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-12B  
**Collection Date:** 11/16/17 5:49

**Lab ID:** 17K0763-24  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:15	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-13A  
**Collection Date:** 11/16/17 5:58

**Lab ID:** 17K0763-25  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:18	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-13B  
**Collection Date:** 11/16/17 5:58

**Lab ID:** 17K0763-26  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:21	EPA200.8 R5.	JTC

**Client Sample ID:** SWF-14A  
**Collection Date:** 11/16/17 6:01

**Lab ID:** 17K0763-27  
**Matrix:** Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:24	EPA200.8 R5.	JTC

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Client: Aires Consulting Group

Project: Berkeley District 87/ Sunnyside School

Lab Order: 17K0763

Client Sample ID: SWF-14B

Lab ID: 17K0763-28

Collection Date: 11/16/17 6:01

Matrix: Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
Metals by ICP-MS									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:26	EPA200.8 R5.	JTC

Client Sample ID: SWF-15A

Lab ID: 17K0763-29

Collection Date: 11/16/17 6:02

Matrix: Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
Metals by ICP-MS									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:29	EPA200.8 R5.	JTC

Client Sample ID: SWF-15B

Lab ID: 17K0763-30

Collection Date: 11/16/17 6:02

Matrix: Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
Metals by ICP-MS									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:32	EPA200.8 R5.	JTC

Client Sample ID: SWF-16A

Lab ID: 17K0763-31

Collection Date: 11/16/17 6:06

Matrix: Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
Metals by ICP-MS									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:35	EPA200.8 R5.	JTC

Client Sample ID: SWF-16B

Lab ID: 17K0763-32

Collection Date: 11/16/17 6:06

Matrix: Drinking Water

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
Metals by ICP-MS									
*Lead	U	2.00		µg/L	1	12/7/17 13:29	12/7/17 23:38	EPA200.8 R5.	JTC

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Client: Aires Consulting Group

Project: Berkeley District 87/ Sunnyside School

Lab Order: 17K0763

## Metals by ICP-MS - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC %REC	%REC Limits	RPD RPD	RPD Limit	Notes
<b>Batch A007950 - EPA 200.8 Metals</b>										
<b>Blank (A007950-BLK1)</b> Prepared & Analyzed: 12/07/201										
Lead	U	2.00	µg/L							
<b>LCS (A007950-BS1)</b> Prepared & Analyzed: 12/07/201										
Lead	461	2.00	µg/L	500.00		92	85-115			
<b>Matrix Spike (A007950-MS1)</b> Source: 17K0761-09 Prepared & Analyzed: 12/07/201										
Lead	445	2.00	µg/L	500.00	0.188	89	75-125			
<b>Matrix Spike (A007950-MS2)</b> Source: 17K0762-04 Prepared & Analyzed: 12/07/201										
Lead	448	2.00	µg/L	500.00	0.150	90	75-125			
<b>Matrix Spike Dup (A007950-MSD1)</b> Source: 17K0761-09 Prepared & Analyzed: 12/07/201										
Lead	449	2.00	µg/L	500.00	0.188	90	75-125	0.8	20	
<b>Matrix Spike Dup (A007950-MSD2)</b> Source: 17K0762-04 Prepared & Analyzed: 12/07/201										
Lead	452	2.00	µg/L	500.00	0.150	90	75-125	1	20	
<b>Batch A007951 - EPA 200.8 Metals</b>										
<b>Blank (A007951-BLK1)</b> Prepared & Analyzed: 12/07/201										
Lead	U	2.00	µg/L							
<b>LCS (A007951-BS1)</b> Prepared & Analyzed: 12/07/201										
Lead	461	2.00	µg/L	500.00		92	85-115			
<b>Matrix Spike (A007951-MS1)</b> Source: 17K0763-03 Prepared & Analyzed: 12/07/201										
Lead	450	2.00	µg/L	500.00	0.199	90	75-125			

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Client: Aires Consulting Group

Project: Berkeley District 87/ Sunnyside School

Lab Order: 17K0763

## Metals by ICP-MS - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC Limits	RPD Limit	Notes
<b>Batch A007951 - EPA 200.8 Metals</b>								
<b>Matrix Spike (A007951-MS2)</b>		<b>Source: 17K0763-13</b>			<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	441	2.00	µg/L	500.00	0.305	88	75-125	
<b>Matrix Spike Dup (A007951-MSD1)</b>		<b>Source: 17K0763-03</b>			<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	454	2.00	µg/L	500.00	0.199	91	75-125	0.9 20
<b>Matrix Spike Dup (A007951-MSD2)</b>		<b>Source: 17K0763-13</b>			<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	448	2.00	µg/L	500.00	0.305	90	75-125	2 20
<b>Batch A007952 - EPA 200.8 Metals</b>								
<b>Blank (A007952-BLK1)</b>					<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	U	2.00	µg/L					
<b>LCS (A007952-BS1)</b>					<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	458	2.00	µg/L	500.00		92	85-115	
<b>Matrix Spike (A007952-MS1)</b>		<b>Source: 17K0763-23</b>			<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	451	2.00	µg/L	500.00	0.186	90	75-125	
<b>Matrix Spike Dup (A007952-MSD1)</b>		<b>Source: 17K0763-23</b>			<b>Prepared &amp; Analyzed: 12/07/201</b>			
Lead	428	2.00	µg/L	500.00	0.186	86	75-125	5 20

Prairie Analytical Systems, Inc.

Date: 12/12/2017

---

---

**LABORATORY RESULTS**

**Client:** Aires Consulting Group

**Project:** Berkeley District 87/ Sunnyside School

**Lab Order:** 17K0763

---

**Notes and Definitions**

- \* NELAC certified compound.
- U Analyte not detected (i.e. less than RL or MDL).

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

<b>Client:</b>	Aires Consulting Group					<b>Lab Order:</b>	17K0766			
<b>Project:</b>	Berkeley District 87/ Sunnyside School					<b>Lab ID:</b>	17K0766-01			
<b>Client Sample ID:</b>	SWF-17A					<b>Matrix:</b>	Drinking Water			
<b>Collection Date:</b>	11/21/17 7:10									
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:40	12/8/17 10:56	EPA200.8 R5.	LAH	
<b>Client Sample ID:</b>	SWF-17B					<b>Lab ID:</b>	17K0766-02			
<b>Collection Date:</b>	11/21/17 7:10					<b>Matrix:</b>	Drinking Water			
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:11	EPA200.8 R5.	LAH	
<b>Client Sample ID:</b>	SWF-18A					<b>Lab ID:</b>	17K0766-03			
<b>Collection Date:</b>	11/21/17 7:07					<b>Matrix:</b>	Drinking Water			
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:18	EPA200.8 R5.	LAH	
<b>Client Sample ID:</b>	SWF-18B					<b>Lab ID:</b>	17K0766-04			
<b>Collection Date:</b>	11/21/17 7:07					<b>Matrix:</b>	Drinking Water			
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:20	EPA200.8 R5.	LAH	
<b>Client Sample ID:</b>	SWF-19A					<b>Lab ID:</b>	17K0766-05			
<b>Collection Date:</b>	11/21/17 7:05					<b>Matrix:</b>	Drinking Water			
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:22	EPA200.8 R5.	LAH	
<b>Client Sample ID:</b>	SWF-19B					<b>Lab ID:</b>	17K0766-06			
<b>Collection Date:</b>	11/21/17 7:05					<b>Matrix:</b>	Drinking Water			
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:25	EPA200.8 R5.	LAH	
<b>Client Sample ID:</b>	SWF-20A					<b>Lab ID:</b>	17K0766-07			
<b>Collection Date:</b>	11/21/17 7:02					<b>Matrix:</b>	Drinking Water			
<b>Analyses</b>	<b>Result</b>	<b>Limit</b>	<b>Qual</b>	<b>Units</b>	<b>DF</b>	<b>Date Prepared</b>	<b>Date Analyzed</b>	<b>Method</b>	<b>Analyst</b>	
<b>Metals by ICP-MS</b>										
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:27	EPA200.8 R5.	LAH	

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Analyses	Result	Limit	Qual	Units	DF	Date Prepared	Date Analyzed	Method	Analyst
<b>Client:</b> Aires Consulting Group									
<b>Project:</b> Berkeley District 87/ Sunnyside School									
<b>Lab Order:</b> 17K0766									
<b>Client Sample ID:</b> SWF-20B									
<b>Collection Date:</b> 11/21/17 7:02									
<b>Lab ID:</b> 17K0766-08									
<b>Matrix:</b> Drinking Water									
<b>Analyses</b>									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:29	EPA200.8 R5.	LAH
<b>Client Sample ID:</b> SWF-21A									
<b>Collection Date:</b> 11/21/17 6:59									
<b>Lab ID:</b> 17K0766-09									
<b>Matrix:</b> Drinking Water									
<b>Analyses</b>									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:38	EPA200.8 R5.	LAH
<b>Client Sample ID:</b> SWF-21B									
<b>Collection Date:</b> 11/21/17 6:59									
<b>Lab ID:</b> 17K0766-10									
<b>Matrix:</b> Drinking Water									
<b>Analyses</b>									
<b>Metals by ICP-MS</b>									
*Lead	U	2.00		µg/L	1	12/8/17 7:42	12/8/17 11:40	EPA200.8 R5.	LAH

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Client: Aires Consulting Group

Project: Berkeley District 87/ Sunnyside School

Lab Order: 17K0766

## Metals by ICP-MS - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
<b>Batch A007967 - EPA 200.8 Metals</b>										
<b>Blank (A007967-BLK1)</b>										
Prepared & Analyzed: 12/08/201										
Lead	U	2.00	µg/L							
<b>LCS (A007967-BS1)</b>										
Prepared & Analyzed: 12/08/201										
Lead	442	2.00	µg/L	500.00		88	85-115			
<b>Matrix Spike (A007967-MS1)</b>										
Source: 17K0764-01 Prepared & Analyzed: 12/08/201										
Lead	480	2.00	µg/L	500.00	0.266	96	75-125			
<b>Matrix Spike (A007967-MS2)</b>										
Source: 17K0764-11 Prepared & Analyzed: 12/08/201										
Lead	483	2.00	µg/L	500.00	1.67	96	75-125			
<b>Matrix Spike Dup (A007967-MSD1)</b>										
Source: 17K0764-01 Prepared & Analyzed: 12/08/201										
Lead	558	2.00	µg/L	500.00	0.266	111	75-125	15	20	
<b>Matrix Spike Dup (A007967-MSD2)</b>										
Source: 17K0764-11 Prepared & Analyzed: 12/08/201										
Lead	467	2.00	µg/L	500.00	1.67	93	75-125	3	20	
<b>Batch A007968 - EPA 200.8 Metals</b>										
<b>Blank (A007968-BLK1)</b>										
Prepared & Analyzed: 12/08/201										
Lead	U	2.00	µg/L							
<b>LCS (A007968-BS1)</b>										
Prepared & Analyzed: 12/08/201										
Lead	442	2.00	µg/L	500.00		88	85-115			
<b>Matrix Spike (A007968-MS1)</b>										
Source: 17K0766-02 Prepared & Analyzed: 12/08/201										
Lead	543	2.00	µg/L	500.00	0.0267	109	75-125			

Prairie Analytical Systems, Inc.

Date: 12/12/2017

## LABORATORY RESULTS

Client: Aires Consulting Group

Project: Berkeley District 87/ Sunnyside School

Lab Order: 17K0766

## Metals by ICP-MS - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	--------------------	-------	----------------	------------------	------	----------------	-----	--------------	-------

## Batch A007968 - EPA 200.8 Metals

Matrix Spike (A007968-MS2)		Source: 17K0767-02			Prepared & Analyzed: 12/08/201					
Lead	501	2.00	µg/L	500.00	0.568	100	75-125			
Matrix Spike Dup (A007968-MSD1)		Source: 17K0766-02			Prepared & Analyzed: 12/08/201					
Lead	533	2.00	µg/L	500.00	0.0267	107	75-125	2	20	
Matrix Spike Dup (A007968-MSD2)		Source: 17K0767-02			Prepared & Analyzed: 12/08/201					
Lead	477	2.00	µg/L	500.00	0.568	95	75-125	5	20	

Prairie Analytical Systems, Inc.

Date: 12/12/2017

---

---

**LABORATORY RESULTS**

**Client:** Aires Consulting Group

**Project:** Berkeley District 87/ Sunnyside School

**Lab Order:** 17K0766

---

**Notes and Definitions**

- \* NELAC certified compound.
- U Analyte not detected (i.e. less than RL or MDL).